



Corporate Office:

1150 Simpson Way
Escondido
California
92029
760-466-1060
760-466-1057 fax
www.QuakeHOLDIndustrial.com

Los Angeles Office:


411 S. Vermont Avenue
Glendora
California
91741
626-852-8800
626-852-8801 fax

San Francisco Bay Area

Office:

22113 Queen Street
Castro Valley
California
94546
800-959-4053
510-363-8067 fax

Freezer/Incubator Wall Mount to Strut



Name: Freezer/Incubator Wall Mount to Strut

Description: Strut attached to wall studs - Equipment attached to strut at 4 locations

Part Number: U4.S16.G (2)

Color: Gray

Kit Contains: 2 rows of steel strut - 4 safety straps with Industrial Velcro

Fastening Method: Steel strut secured to wall studs - Safety straps secured to equipment with Industrial Velcro

Strength Ratings: Each safety strap is rated at 792 pounds shear strength